

## CueSee® Hypoxic

### X-Low

**LOT** 16301812



**REF** 166.001.005

### Radiometer ABL 835 FLEX

Analyte	Mean	Range	Unit
$pO_2$	20.6	8.6 – 32.6	mmHg

### Abbott i-STAT® system CLEW: A38

Cartridge types: CG8+, EG7+, EG6+, G3+ & CG4+ Lots: A, K, L, M, N, W, Y

Analyte	Mean	Range	Unit
$pO_2$	18	6 – 30	mmHg

Cartridge types: G3+ & CG4+ Lot: D

Analyte	Mean	Range	Unit
$pO_2$	17	5 – 29	mmHg

### epoc® Blood Analysis System

See instrument manufacturer's website for Assigned values

Assigned values are established on the instruments listed. For other instruments mean values and ranges can be established by the end user.

AN01183A01 – 20180615